

Office of the Director
DR. BHUBENESWAR BOROOAH CANCER INSTITUTE
A grant-in-aid institute of Department of Atomic Energy, Govt. of India
And a unit of Tata Memorial Centre (Mumbai)
Gopinath Nagar, Guwahati – 781016

No. BBCI-TMC/GeM/B/5950204/657 /2025

Dated: 25.02.2025


CORRIGENDUM

GEM NO: GEM/2025/B/5950204

Tender Date: 14-02-2025

Corrigendum No: 01

Sr. No	BBCI Technical Specification	Corrigendum
7	Automated dilution of samples and controls with the ability to do automatic re- dilution in case of test results beyond the linearity of the assay (rerun) and reflex testing. Have the possibility to detect factor inhibitors by performing automated multiple dilutions (factor parallelism). The ability to show the reaction curves (Clot Curves) of the patient results.	Automated dilution of samples and controls with the ability to do automatic re- dilution in case of test results beyond the linearity of the assay (rerun) and reflex testing. Have the possibility to detect factor inhibitors. The ability to show the reaction curves (Clot Curves) of the patient results.



Director

Dr. B. Borooah Cancer Institute