



DR. BHUBANESWAR BOROOAH CANCER INSTITUTE
PROCUREMENT DEPARTMENT

ITEM DESCRIPTION : CORRIGENDUM FOR “LOW TEMPERATURE HYDROGEN PEROXIDE GAS
STERILIZER (PLASMA STERILIZER)”
GEM BID NO : GEM/2024/B/5062089

Sr. No	Original Technical Specification	Revised Technical Specification
1.	STERILIZER, READER(with printer) and BI Indicator should be from same manufacturer.	Read as: STERILIZER, READER (with printer) and BI Indicator should be from same manufacturer or compatible manufacturer of reputed manufacturer.
2.	Lumen sterilization claims should be validated & endorsed by USFDA only.	Rad as: Lumen sterilization claims should be validated & endorsed by USFDA/ European CE (04 digit NB). (It must be authenticated through their official website ,third party certification will not be considered)
3.	Should have local service facility from OEM company based in East India & submit details (Escalation Matrix) in the technical Bid.	Read as : Should have local service facility from OEM / Authorized Dealer company based in East India & submit details (Escalation Matrix) in the technical Bid.

BL/19 9/7/24

Chief Administrative Officer
Dr. B. Borooah Cancer Institute, Guwahati