

DR. B. BOROOAH CANCER INSTITUTE
PROCUREMENT DEPARTMENT

Subject : Corrigendum-I
GEM Bid No : GEM/2025/B/6116025
Item Description : Washer Disinfector (Single Door)

SI No	Tender Specification	Revised Specification
4	Technical Specs Point no. 28.vi: vi. Standardize Detergent for washer— 6 cans	Revised as- Standardize Detergent from own manufacturing or recommended by the manufacturer to ensure efficiency and longevity of the equipment along with a list of recommended detergents by the manufacturer should be submitted. No. of detergent- 6 cans.
7	Technical Specs Point no. 30.i): i) The equipment should conform to international safety & quality standards with proper certifications such as European CE (four-digit notified body) / US-FDA / equivalent BIS for quoted model and CDSCO.	Revised as- The Steam Sterilizer should conform to international safety & quality standards with proper certifications such as a. European CE (four-digit notified body) or b. US-FDA or c. CDSCO Class-C with ISO 14937:2009, ISO 11134:1994, ISO 9001:2015, ISO 15883 [All ISO must be issued by IAF Accredited Agency]
11	Technical Specs Point no. 31.ii): ii). There should be a minimum of 06 (Six) installations with performance certificates for the quoted model in India, preferably from reputed government hospitals such as AIIMS, JIPMER, TMC, PGI Chandigarh, etc. The OEM or their subsidiary or authorized dealer should be in India, having similar Washer Disinfector selling experience of more than 5 years. Performance/ Satisfactory installation reports should be enclosed with the bid.	Revised as- The quoted model must have at least 6 (six) installations in India, each accompanied by performance certificates, in government hospitals or NABH-accredited private hospitals with 300 or more beds.

*All other specification and terms & conditions of GEM/2025/B/6116025 will remain unchanged.

Competent Authority