

DR. BHUBANESWAR BOROOAH CANCER INSTITUTE

PROCUREMENT DEPARTMENT

ITEM DESCRIPTION

ADDENDUM FOR "VOLUMETRIC INFUSION PUMP"

GEM BID NO

GEM/2024/B/5107508

Sr. No	Revised Technical Specification
1.	Should have following Infusion Modes: mL/h modes: Volume + Flow rate, Volume + Time or Equivalent Modes.
2.	IV lines should be DEHP-free & Latex-free, PVC-free.
3.	The infusion pump should be supplied with individual clamp & power cord for each unit, and should have compatibility with docking system for future upgrades.

Director

Dr. B. Borooah Cancer Institute