

Dr. Bhubaneswar Borooh Cancer Institute

A Grant-in-Aid Institute of Department of Atomic Energy, Govt. of India And a Unit of Tata Memorial Centre (Mumbai) Gopinath Nagar, A. K. Azad Road, Guwahati-781016, Assam, India

No. BBCI-TMC/Misc-216/Store/1084/2020

Date:

NOTICE INVITING QUOTATION

Dr. B. Borooh Cancer Institute, Gopinath Nagar, Guwahati-16 invites quotations from the reputed company, manufacturer, authorized supplier / dealer etc. for supply & installation of **3 para monitor with stand** (estimated qty. – 14 Nos.) as per technical specification mentioned in Annexure A.

Terms & conditions:

1. The quantity mentioned in the Notice Inviting Quotation of the above items is an estimate only. Actual requirement of the items may vary (less or more) from the estimated quantity which will be decided by the Management of Dr. B. Borooh Cancer Institute at the time of purchase.
2. The quotation must accompany the detailed technical specifications, user manuals and certificates of the quoted products.
3. All information asked for must be provided in the same order as in the specifications. Incomplete and ambiguous information will not be accepted.
4. Please mention the make / brand & model or product quote of the item.
5. The valid technical offers will be evaluated by a Technical Assessment Committee (TAC) of the Institute.

The TAC will recommend the technically short-listed vendors for further consideration.

The offers of the technically shortlisted vendors / products will only be considered for price comparison.

6. The successful quotationer shall have to supply & install the equipment to this Institute at his own risk and responsibility and in good condition in all respects.
After successful installation expertise from the supplier should give proper demonstration to the professional staff of the Institute for the use of the equipments.
7. The standard warranty conditions of the quoted products should be as per warranty term mentioned in the respective Annexure
8. **An Earnest Money Deposit of Rs. 14000/- (Rupees fourteen thousand only) is to be paid by DD in favour of Dr. Bhubaneswar Borooh Cancer Institute and should be enclosed in the Part-I Technical Bid of the tender documents. The EMD DD number and date should be mentioned on the envelope.**

(contd..2/p.)

9. The quotation notice is to be duly signed on all the pages as a token of having accepted all the terms and conditions (no thumb impression is to be affixed).
10. **The quotation shall consist of 2 parts, namely Part-I (Technical) and Part-II (Commercial), in separate sealed covers. Part-I & Part-II offers shall be submitted in one sealed cover and this will be received up to 2:00 PM of 24.03.2019 and the same will be opened on the same day at 3:00 PM.**
11. The quotationer should submit their relevant credentials along with the Technical Offer (Part-I) of the Quotation as follows:
- (a) Particulars of the participating firm / vendor / company (quotationer)
 - (b) Certificate of previous experience of supply & installation of similar product with list of institutions (in case of supplier); OR,
 - (c) Certificate of dealership / distributor etc. from the appropriate manufacturer in case the quotationer is a dealer / distributor; OR,
 - (d) Certificate in regard to license of manufacturing in case of manufacturer.
 - (e) Certificate in regard to registration of firms along with trade license
- Vendors are advised to make compliance statement of technical specifications of the quoted products with make and model as per mentioned in the respective
- (f) Annexure.
 - (g) Bank Account details of the company / vendor
 - (h) GST Registration Number
 - (i) PAN Number
12. The price should be quoted in the following format which should be inclusive of all charges. Taxes, if any, should be shown separately. However, GST will be paid to the successful vendor on prevailing rate at the time of billing / payment.

Sl.	Item	Unit price (Rs.)	Qty.	Amount before GST (Rs.)	GST %	Any other taxes, if applicable	Grand Total (after GST) : (Rs.)

13. The quotation shall include delivery time, warranty, and payment terms, etc and other relevant terms & conditions.
14. The rate should be quoted both in figure and in words.
15. The rate if accepted will remain valid for 1(one) year from the date of acceptance and there will no change under any circumstances during the period.
16. Payment will be made within 60 (sixty) days of submission of bill after successful supply & installation of the items after observing the requisite formalities.
17. Incomplete quotation not in conformity with the terms and conditions as in this quotation notice will be rejected outright.
18. The undersigned reserves the right either to accept or to reject any or all the quotations without assigning any reason thereof and is not bound to accept the lowest quotation.
19. Decision of the Technical Assessment Committee / Purchase Committee / any other committee constituted by the authority of Dr B Borooah Cancer Institute shall be final and binding in all cases.



Director

Dr B Borooah Cancer Institute
Guwahati-781016

To,

- (1) BBCI Website www.bbcionline.org
- (2) BBCI Notice Board
- (3) Various Guwahati based dealers / suppliers

ANNEXURE- A

Item	Technical Specification:
3 -para Patient Monitor	1. Should have ECG, SpO2 and NIBP as standard parameters.
	2. Should display at least 3 waveforms of selected parameters simultaneously.
	3. Should have inbuilt continuous battery backup of minimum two hours.
	4. Display: Color TFT display size of 10" or more.
	5. Should have facility for displaying multi-screen configurations
	6. Should be able to store & display at least 72 hours of tabular & graphical trends of all parameters. Should also have full disclosure of waveforms for minimum 48 hours.
	7. Should be suitable for monitoring pediatric and adult patients.
	8. The SpO2 technology should be Fast & should sense in hypotension, shivering & motion
	9. Should have different patient type selection
	10. Should be able to analyze arrhythmias & ST segment changes.
	11. Should be able to give visual & audible alarms with three levels of volume adjustment.
	12. Should have connectivity to Central station through Ethernet card or Wireless Connectivity.
	13. USB port should be available to directly export tabular trend data. Monitor should also be able to export data in HL7 format.
	14. Should be European CE or US-FDA approved.
	15. Scope of supply: Equipment should be supplied with compatible a) 5 lead ECG cable (1 No.), b) SpO2 (Pediatric and Adult) probe 1 No. each, c) Connecting hose-tube with NIBP Cuff Adult & Pediatric (1 each) d) Operating/user Manual (1 No.)
	16. Warranty: 2 years on equipment from the date of installation along with all accessories. CMC for 8 years after completion of warranty period of 2 years.
	17. The company should mention the make & model name/number of the quoted equipment and submit the technical brochure of the quoted model in the Technical Bid along with compliance sheet as per technical specifications.
	18. The unit shall be capable of stored and operating continuously in ambient temperature of 10 - 50deg C and relative humidity of 15-90% .
	19. User Manual and service manual in English should be handed at the time of installation.
	20. Should have local service facility .The service provider should have the necessary equipments recommended by the manufacturer to carry out preventive maintenance test as per guidelines provided in the service/maintenance manual.
	21. List of important spare parts and accessories with their part number and costing should be given in commercial offer.
22. Power Supply: Power input to be 220-240VAC, 50Hz fitted with Indian plug.	
23. Post Installation end user training is must.	