

DR. BHUBENESWAR BOROOAH CANCER INSTITUTE

A grant-in-aid institute of Department of Atomic Energy, Govt. of India

And a unit of Tata Memorial Centre (Mumbai)

Gopinath Nagar, Guwahati – 781016

No. BBCI-TMC/GeM/B/5888822/450 /2025

Dated: 11/02/2025

CORRIGENDUM

GEM NO: GEM/2025/B/5888822

Tender Date: 30-01-2025

Corrigendum No: 01

Sr. No.	BBCI Technical Specification	Revised Specification
i	One Desktop Computer of reputed brand with following Specification to be supplied: Intel Core i7 of latest generation or better processor. Hard disk: 1 TB - 4 TB. Graphic card 2GB RAM: 16GB or better. Operating System of Window 10 (64 bits) or latest multimedia. DVD ROM + writer. 2 Ethernet interfaces. Display: At least 27" or better LED IPS Monitor with resolution of 2560x1440 p. - 2(two) Nos.	One Desktop Computer of reputed brand with following Specification to be supplied: Intel Core i7 of latest generation or better processor. Hard disk: 1 TB - 4 TB. Graphic card 2GB RAM: 16GB or better. Operating System of Window 10 (64 bits) or latest multimedia. DVD ROM + writer. 2 Ethernet interfaces. Display: At least 24" or better LED IPS Monitor with resolution of At least 1920*1080p. - 1(one) Nos.
ii	3KVA Online UPS with minimum 15 minutes back-up.	3KVA Online UPS with isolation transformer with minimum 15 minutes back-up (Vertive/APC/Riello)
iii	3 KV or suitable Servo stabilizer Compatible with the OPG System with spike suppressor to be supplied.	Omitted
vi	Vendor should also supply 5 (Five) Nos. of packet (500) Films of size 8" x 10" compatible with the existing DICOM printer.	Vendor should also supply 5 (Five) Nos. of packet (500) Films of size 8" x 10" compatible with the existing AGFA DICOM printer.
vii	Should be supplied with 01 (one) No. of office table with drawers of suitable size to accommodate computer system and 2 (two) Nos. of height- adjustable chairs (Godrej/Fetherlite make).	Should be supplied with 01 (one) No. of office table (Size: 4*2.5) with drawers of suitable size to accommodate computer system and 2 (two) Nos. of height- adjustable chairs (Godrej/Fetherlite make).
7	Should have an option to Upgrade to 3D on a later date (The Cost of the upgradation should be quoted separately).	May preferably have an option to Upgrade to 3D on a later date (The Cost of the upgradation should be quoted separately) (Optional).

13/02/25
Chief Administrative Officer
 Dr. B. Borooah Cancer Institute