

**CORRIGENDUM**

Sub: Corrigendum for Video Stroboscope with FEES system

GeM Bid No: GEM/2024/B/4489011, Dated: 25-01-2024

Sl. No. for Tender Bid	Technical Specifications of tender	Amendment Requested	Revised Technical Specifications
<b>B.</b>	<b>Digital Colour High Definition Video Processor:</b>		
2	Should be provided <i>with an integrated or separate 300 W Xenon light source/LED</i>	"Should be provided with <b>all main units like Camera and light Source separate</b> ".	Already included no change
3	Should have Optical or Digital Image Enhancement of capillary vessels and fine patterns in the superficial	"Should have <b>Optical or Digital Image enhancement of capillary vessels and fine patterns in the superficial (Optional)</b> ".	No change
4	Layer of mucosa for early detection of lesions	"Layer of mucosa for early detection of lesions <b>(Optionals)</b> ".	No change
10	Colour balance: Red/Blue levels selectable	"Colour balance: Red/Blue levels selectable <b>(Optional)</b> ".	Ok
14	Should have the facility for Transnasal Esophagoscopy (TNE) with pump mode.	"Should have the facility for Transnasal Esophagoscopy (TNE) with pump mode <b>(Optional)</b> "	No change
<b>C.</b>	<b>Video Naso-Pharyngo-Laryngoscope (Pediatric) Quantity 1 No.:</b>		
2	Distal <i>end</i> Diameter should be 2.2 to 2.6 mm or less.	"Distal <b>tip outer</b> diameter should be <b>2.2 to 2.9 mm</b> ".	Ok
3	Insertion tube outer diameter should be 2.9 mm or less.	"Insertion tube outer diameter should be <b>3 mm</b> or less".	Ok
6	Angulation range should be approx. 130° UP & 130° DOWN.	Angulation should be <b>in the range of 130° - 140° upwards and downwards</b> ".	Included in Spec. No change
<b>D.</b>	<b>Video Naso-Pharyngo-Laryngoscope (Adult) Quantity 2 Nos.:</b>		
2	Distal end Diameter should be 2.6 mm less.	"Distal <b>tip outer</b> diameter should be <b>3.7 mm or less</b> ".	No change
3	Insertion tube outer diameter should be 3.3 mm or less.	"Insertion tube outer diameter should be <b>4 mm</b> or less".	No change
6	Angulation range should be approx. 130° UP & 130° DOWN.	Angulation should be <b>in the range of 130° - 140° upwards and downwards</b> ".	Included. No change

## DR. BHUBANESWAR BOROOAH CANCER INSTITUTE

A grant-in-aid institute of Department of Atomic Energy, Govt. of India

And a unit of Tata Memorial Centre (Mumbai)

Gopinath Nagar, Guwahati- 781016

\*\*\*\*\*

E.	<b>Digital HD Endoscopy recording System:</b>	<b>"Medical Grade Inbuilt/External HD Recording System with Patient Data Entry:"</b>	No change
6	Internal HDD capacity should be at-least 4 TB with portal and server.	"Internal/External HDD capacity should be at-least 4 TB".	Ok
7	Should have functions for editing of still images & videos like cutting, cropping etc.	"Kindly <u>delete/remove</u> this point completely from the existing prescribed technical Specifications".	No change
8	Should have functions to play the video in slow motion, audio visual format.	"Please <u>erase/eliminate</u> the said entire point from the proposed Technical Specifications".	No change
9	Should have integrated file viewer for still images & videos.	"Kindly <u>exclude/discard</u> this point from the given Technical Specifications".	No change
10	Should have structured & clear user guidance to fill patient's data (demographic details), and have access for functions like play, recording start-stop, pause, capture.	"Please <u>drop/extract</u> the said point from the proposed Technical Specifications".	No change
12	Should have an in built display monitor or separate HD monitor should be provided which allows to record or capture in real time.	"Kindly <u>remove/eliminate</u> this point completely from the existing prescribed technical Specifications".	No change
Further, request you to kindly add and incorporate the following additional point at the end of the existing Technical Specifications under the <b>Heading - E : "Digital HD Endoscopy Recording System"</b> vide aforesaid item:			
<b>Should be supplied with a laptop with patient reporting software</b> (It will be more convenient for the doctors to generate report and analyze the case.)			No change




Tel: 09957033199 / 09957033212

Office of the Director

**DR. BHUBANESWAR BOROOAH CANCER INSTITUTE**

A grant-in-aid institute of Department of Atomic Energy, Govt. of India

And a unit of Tata Memorial Centre (Mumbai)

Gopinath Nagar, Guwahati- 781016

\*\*\*\*\*

Dated: 08/03/2024

**TIME EXTENSION NOTICE FOR SUBMISSION OF BID**

Due to technical reason, the last date for online submission of bid for supply, installation & commissioning of **Video Stroboscope with FEES System** against **GeM Bid No: GEM/2024/B/4489011, dtd. 25-01-2024** at Dr. B Borooah Cancer Institute has been extended upto **02:00 PM of 15-03-2024**. Bid will be open on **15-03-2024 at 02:30 PM**. Along with this a corrigendum has been uploaded for the respective bids. For detail corrigendum, Visit GeM website [www.gem.gov.in](http://www.gem.gov.in) or BBCI website [www.bbcionline.org](http://www.bbcionline.org). Other terms & conditions of the following tenders shall remain unchanged.

*13/19 8/3/24*

Chief Administrative Officer  
Dr. B. Borooah Cancer Institute  
*dh* Guwahati - 16

No. BBCI-TMC/GeM/B/ 920 /2024

Dated: 08 /03 /2024

Copy for information to:

1. BBCI Website
2. Notice Board, BBCI
3. Concerned file

*13/19 8/3/24*

Chief Administrative Officer  
Dr. B. Borooah Cancer Institute  
*dh* Guwahati - 16