

Office of the Director  
**Dr. Bhubaneswar Borooah Cancer Institute**  
A Grant-in-Aid Institute of Department of Atomic Energy, Govt. of India  
and a Unit of Tata Memorial Centre (Mumbai)  
Gopinath Nagar, Guwahati- 781016

No. BBCI-TMC/Equip-446/ 433 /2022

Date: 31.03.2022

LIMITED TENDER

Dr. B. Borooah Cancer Institute, Gopinath Nagar, Guwahati- 781016 invites quotations from authorized vendors to provide offers of the following service:

Sl. No	Particulars	Quantity	Estimated Value (in Rs.)
1.	Video Recording System	01 no	Rs. 4,00,000.00

1. The quotation must be in the same order as in the specifications and should be free from Correction / erasures. In case there is any unavoidable correction it should be properly attached. If not, the quotation will not be considered. Quotations written in pencil will not be considered.
2. The vendor should provide authorization letter along with their offer.
3. The quotation shall be submitted in one sealed form at Dr. BBCI office and this will be received up to **2:00 PM of 21.02.2022** and the same will be opened on the same day at **3:00 PM**.
4. No tender will be accepted unless the full amount of the Earnest Money Deposit of **Rs.8000/-** (Rupees Three thousand only) is paid. EMD is to be paid **by DD in favour of Dr. Bhubaneswar Borooah Cancer Institute**. Original receipt must be enclosed with the tender document.
5. The quotation must accompany the Product details.
6. Director, Dr. B. Borooah Cancer Institute reserves the right to accept the offer by individual terms and reject any or all tenders without assigning any reason thereof and does not bind itself to accept lowest quotations.
7. Please mention the make/brand and model of the quoted product. Also submit necessary leaflet of the quoted product.
8. The quotation notice number must be clearly mentioned on the envelope.
9. The valid technical offers will be evaluated by a Technical Assessment Committee (TAC) of the Institute. The TAC will recommend the technically short listed vendors for further consideration. The offers of the technically short listed vendors/products will only be considered for price comparison.
10. The successful quotationer shall have to supply the materials to this institute at his own risk and responsibility and in good condition in all respects.
11. The quotation notice is to be duly signed on all pages as a token of having accepted all terms and conditions (no thumb impression is to be affixed).

12. The price should be quoted in the following format which should be inclusive of all charges. Taxes, if any, should be shown separately. However, GST will be paid to the successful vendor at prevailing rate at the time of billing/ payment.

SL No	Item	Unit Price(Rs.)	Qty.	Amount before GST(RS.)	GST%	Any other taxes, if applicable	Grand Total (after GST) (Rs.)

13. The quotation shall include delivery time, payment terms, warranty etc. and other relevant T & C.
14. The rate should be quoted both in figure and in words.
15. Payment terms: Payment will be made within 60 (sixty) days of submission of bill after successful delivery of the items.
16. Incomplete quotation not in conformity with the terms and conditions as in this quotation notice will be rejected outright.
17. Decisions of the Technical Assessment Committee/ Purchase Committee/ any other committee constituted by the authority of Dr. B. Borooah Cancer Institute shall be final and binding in all cases.
18. The tender shall be valid for a period of 180 days from the date of the opening of the bid. Validity of the tender shall be extended if required.

*132-19 31/01/22.*  
**Sr. Administrative Officer**  
Dr. B. Borooah Cancer Institute  
Guwahati- 78 1016

Copy to:

1. BBCI website [www.bbcionline.org](http://www.bbcionline.org)
2. BBCI Notice Board

## Technical Specification

<b>HD Recorder Cum Digital Documentation System for High Definition Recording &amp; Data Archiving</b>	
<b>Sl. No</b>	<b>Specifications</b>
The Full High-Definition Digital Documentation System for recording, managing, and archiving surgical images and video in native full HD resolution. The captured full high definition images & videos should be accessible from the hard drive for printing or saving onto multiple forms of external media which includes:-	
1	Full High Definition Digital Documentation System (Recorder System)
2	Should have 3.5" – 4" LCD Monitor for Live view, play back or GUI
3	Multiple HD & SD inputs/ outputs (SDI, DVI-D, S-VIDEO and composite)
4	Simultaneous recording to multiple media (upto 2 external USB drives & 1 internal HD)
5	Networking (FTP & CIFS) RJ45 10/ 100/1000 Ethernet
6	Should be real time, MPEG 2 HD MPEG4, 1 or 2 compression engine with full IBP encoding should have inputs minimum 2 nos. S-Video, 2 nos. Composite, 1 XGA (1920 x 1080p)
7	Image capture:1080p,720p, 1280x1024p, (PAL). Should have Stereo audio inputs
8	Should have internal Disc Capacity of 500GB or more
9	The recorder should be compatible to control remotely from existing Stryker 1488 camera head should have the following options:-
a)	Camera Head Triggering Remote Control.
b)	Patient Information and Device Management inbuilt on screen
c)	On-console LCD Control.
d)	Should be US FDA Approved.