

Tel: 09957033199 / 09957033212

Office of the Director

DR. BHUBANESWAR BOROAH CANCER INSTITUTE

A grant-in-aid institute of Department of Atomic Energy, Govt. of India

And a unit of Tata Memorial Centre (Mumbai)

Gopinath Nagar, Guwahati- 781016

No- BBCI-TMC/Accts/334/ *48A* /2022

Dated: *04/02/2022*

CORRIGENDUM

The technical specification for the supply & installation of **Shaking Water Bath (Qty- 01 nos)** at Dr. B. Borooah Cancer Institute against **NIT NO.: BBCI-TMC/ Accts/334/ 198 / 2022 dated 17/01/2022** has been revised. The revised technical specification for the above mentioned equipment shall be read as follows:

Sl. No	Equipment	As Published in tender document	Ammendment
1	Shaking Water Bath	Temperature working range should be from amb. +7 to +100 degC or better range. Sl. No J – (8)	Temperature working range should be within ambient 10°C to 100°C

The last date of submission of the tender has been extended till **10.02.2022 up to 2:00 PM** and the same will be opened on the same day at **3:00 PM**.

The other specifications and Terms & Conditions of the NIT shall remain unchanged.

B L 19 4/2/22
Senior Administrative Officer
Dr. B. Borooah Cancer Institute
Guwahati – 16

Copy for information to:

1. BBCI website www.bbcionline.org
2. Notice Board, BBCI
3. Concerned file

B L 19 4/2/22
Senior Administrative Officer
Dr. B. Borooah Cancer Institute
Guwahati – 16