

Office of the Director
Dr. Bhubaneswar Borooah Cancer Institute
A Grant-in-Aid Institute of Department of Atomic Energy, Govt. of India
and a Unit of Tata Memorial Centre (Mumbai)
Gopinath Nagar, Guwahati- 781016

No. BBCI-TMC/Accts/334/ 198 /2022

Date: 17 / 01 / 2022

LIMITED-TENDER NOTICE

Dr. B. Borooah Cancer Institute, Gopinath Nagar, Guwahati- 781016 invites quotations from authorized vendors to provide offers of the following service:

Sl. No	Particulars	Quantity	Estimated Value (in Rs.)
1.	Shaking Water Bath	01 no	Rs. 1,50,000.00

1. The quotation must be in the same order as in the specifications and should be free from Correction / erasures. In case there is any unavoidable correction it should be properly attached. If not, the quotation will not be considered. Quotations written in pencil will not be considered.
2. The vendor should provide authorization letter along with their offer.
3. The quotation shall be submitted in one sealed form at Dr. BBCI office and this will be received up to **2:00 PM of 24-01-2022** and the same will be opened on the same day at **3:00 PM**.
4. No tender will be accepted unless the full amount of the Earnest Money Deposit of **Rs.3000/-** (Rupees Three thousand only) is paid. EMD is to be paid **by DD in favour of Dr. Bhubaneswar Borooah Cancer Institute**. Original receipt must be enclosed with the tender document.
5. The quotation must accompany the Product details.
6. Director, Dr. B. Borooah Cancer Institute reserves the right to accept the offer by individual terms and reject any or all tenders without assigning any reason thereof and does not bind itself to accept lowest quotations.
7. Please mention the make/brand and model of the quoted product. Also submit necessary leaflet of the quoted product.
8. The quotation notice number must be clearly mentioned on the envelope.
9. The valid technical offers will be evaluated by a Technical Assessment Committee (TAC) of the Institute. The TAC will recommend the technically short listed vendors for further consideration. The offers of the technically short listed vendors/products will only be considered for price comparison.
10. The successful quotationer shall have to supply the materials to this institute at his own risk and responsibility and in good condition in all respects.
11. The quotation notice is to be duly signed on all pages as a token of having accepted all terms and conditions (no thumb impression is to be affixed).

12. The price should be quoted in the following format which should be inclusive of all charges. Taxes, if any, should be shown separately. However, GST will be paid to the successful vendor at prevailing rate at the time of billing/ payment.

SL No	Item	Unit Price(Rs.)	Qty.	Amount before GST(RS.)	GST %	Any other taxes, if applicable	Grand Total (after GST) (Rs.)

13. The quotation shall include delivery time, payment terms, warranty etc. and other relevant T & C.
14. The rate should be quoted both in figure and in words.
15. Payment terms: Payment will be made within 60 (sixty) days of submission of bill after successful delivery of the items.
16. Incomplete quotation not in conformity with the terms and conditions as in this quotation notice will be rejected outright.
17. Decisions of the Technical Assessment Committee/ Purchase Committee/ any other committee constituted by the authority of Dr. B. Borooah Cancer Institute shall be final and binding in all cases.

BLA 17/01/22
Sr. Administrative Officer
Dr. B. Borooah Cancer Institute
Guwahati- 78 1016

Copy to:

1. BBCI website www.bbcionline.org
2. BBCI Notice Board

Technical Specification

Technical Specification : Shaking Water Bath		
Intended use: Water bath with capacity of 10-25 ltr or more with + 100 deg C heating suitable for saline bottle/ blood bag heating in BMT		
Sl. No	Name of the Vendor:	
A	Technical Specifications/ scope of supply	<u>COMPLIANCE</u> 1) Mention “Complied / Not Complied” 2) Highlight any deviations 3) Mention part number/ Catalogue number of the relevant item quoted.
B	Name of the Manufacturer:	
C	Country of Origin	
D	Model Name / Model No. of the Equipment (if the equipment is a combination of more than one equipment then list down the model no. pertaining to each)	
E	Year of Introduction internationally	
F	Year of introduction in India	
G	Please quote all the models including the latest state of the art model which meets the tender specifications. The offer should be complete in all respect for the intended use.	
H	Technical Specifications / Scope of supply	
I	“Shaking Water Bath for BMT unit-1 No”, complete as per below mentioned configuration and specifications :	
J	Features and specifications required:	
1	Water bath should be rectangular/ square in shape.	
2	Construction should be double walled.	
3	Inner chamber should be made up of rust free stainless steel with seamless corners.	
4	Exterior should be mat finished / powder coated	
5	Gap between the wall should be filled with proper heat insulating materials	
6	Heater- should be corrosion proof and shock proof	
7	Temperature:	
8	Temperature working range should be from amb. +7 to +100 degC or better range	
9	Digital temperature sensor should be available	
10	Temperature setting for Celsius is required.	

11	Microprocessor based heating control	
12	Temperature accuracy error should be +/- 0.1 deg C or less	
13	Digital temperature display – Set and actual temperature display should be available.	
14	Digital timer from 1 min upto 99 hours should be available.	
15	Additional safety thermostat/ Temp. limiter to turn off heater in case of PID failure	
16	Tank- usable water capacity should be 15 to 25 litres.	
17	Tank should have easy cleaning and drainage facility.	
18	Dome shaped/ Gable Lid should be provided with the equipment Lid should be hinged at the back side. Handling for lifting the lid.	
19	Shaker :Shaking facility should be required.	
20	Shaker should be “To and Fro” motion type	
21	Shaker Basket/ tray should be removable	
22	European CE or FDA certified product	
23	Performance parameter required as below: (like accuracy, sensitivity, sample volume requirement, resolution or such performance parameters to be highlighted)	
24	Upgradability capability (List down possible upgrades for the quoted model)	
25	Safety requirements: It should follow International / national safety requirement. Please specify certification with certifying agency and country with copies of certificate.	
26	User’s list: A list of installation with the address and contact numbers to be provided. (User list should be for the quoted model).	
27	<u>Input power supply requirements:</u> 240 V AC ± 10 %, 50 Hz, single phase. Specify if any other power supply requirement is recommended.	
28	After Sales Service Support	
29	<ol style="list-style-type: none"> 1. Company should provide technical support for hardware and software, spares 2. Company should provide required spares for at least 7 years even after discontinuation of the model 3. Bidders who do not qualify this criteria shall not be considered. Bidder should provide address of the authorized service centre and details of numbers of personnel with training credentials. 	

30	Installation, Commissioning, testing and Training	
31	Unpacking and Shifting the consignment to the installation site is to be included in the scope of supply. Bidder / manufacture / authorized service provider should take responsibility to lift/shift the consignment from unloading site to the installation site. Unloading site shall be “Stores Department, KS Building, ACTREC Campus”. If needed, Bidder has to arrange for the labourers at no charge to ACTREC. (Before submitting the quotation, bidders may visit ACTREC to know unloading site and installation site)	
32	Warranty and after sales support	
33	Waranty: 2 years comprehensive-on-site-Factory warranty- for entire equipment.	
34	Free AMC – labour only: 5 years post warranty	
35	Important terms to be noted by the bidders:	
36	Read the above scope of supply carefully and quote accordingly. Incomplete and / or partially complete offers are liable to be rejected.	Agree/ Not Agree
37	Manufacturer / bidder should clearly indicate typical life span / Life expectancy of the quoted equipment. years
38	Mention the time required to install the system. days
39	Bidders should clearly mention about compliance of the warranty and AMC/free service terms in the technical bid. Bids not fully complying to the warranty and AMC/free service terms are liable to be rejected.	Agree/ Not Agree
40	Item/Equipment quoted should not be a second hand/refurbished/on verge of obsolescence /outdated model. Similarly, after award of contract, the item supplied should be a newly manufactured item and not item already manufactured and stocked for more than 1 month.	Agree/ Not Agree
41	After opening of the Technical bid (Part-), Physical demonstration of the quoted model may have to be shown/ arranged by the bidder, if requested by the Institute. Physical demonstration may be shown at one of the end user’s site/ Principle company’s application lab/ manufacturing site located in Mumbai/ Navi Mumbai/ Thane cites. If there are no installations of the quoted model in Mumbai/ Navi Mumbai / Thane cites, then the quoted model may have to be brought in at ACTREC for demo purpose within 10 days from the date of request. Physical Demonstration may be requested to confirm the availability of any or all technical features as mentioned/ stated in the technical bid. Physical Demonstration will also be a part of	Agree/ Not Agree

	technical evaluation process. If the bidder does not comply, such bids are liable to be disqualified. (Demonstration of quoted model is to be shown and not the demonstration of similar models with different technical specifications and features)	
42	Part experience of the bidders in terms of quality of supplied equipment's after sales service and application support will be taken into consideration while technical evaluation. Bidders who has unsatisfactory past experience in last 2-3 years. In terms of quality of supplied equipment's, after sales service and application support, bids of such bidders may liable to be rejected.	Agree/ Not Agree
43	Complete and detailed information should be provided in respect to each point specified in the specifications. Technical bids that are incomplete in any respect are liable to be rejected. Provide relevant supportive information, publications, catalogue, etc. Bidders providing misleading or wrong information are liable to be rejected. All technical claims should be printed in the technical brochure of the equipment.	Agree/ Not Agree
44	If any contradictory statements / figures / information is observed in the compliance chart and in the technical bid, then the technical information mentioned in the product literature / brochure will be considered true and further evaluation will be done based on the information given in the product literature/brochure. Regarding the warranty and free AMC terms, information quoted in the technical bid will only be taken as valid terms.	Agree/ Not agree
45	Compliance column may be filled with relevant data, figures, range etc. as applicable. Do not just mention "YES/ NO/ Complied".	Agree/ Not Agree
46	Declaration by the bidder	
47	We have quoted for all items meeting the description / scope of supply in the Financial bid as per prescribed format of the Tender documents and we agree that Partial/ incomplete offers are liable for rejection.	Yes/No