

Office of the Director  
**DR. BHUBANESWAR BOROOAH CANCER INSTITUTE**

A grant-in-aid institute of Department of Atomic Energy, Govt. of India

And a unit of Tata Memorial Centre (Mumbai)

Gopinath Nagar, Guwahati- 781016

\*\*\*\*\*

No- BBCI-TMC/Equip- 436/2047 /2022

Date: 17.05.2022

**CORRIGENDUM**

The technical specifications for supply, installation & commissioning of “Anesthesia Machine” against tender no: BBCI-TMC/Equip-436/NIT/1723/2022, dtd.26.04.2022 at Dr. B. Borooah Cancer Institute is as follows:

Sl. No	Sl. No of Technical Specification	As published in tender documents	Amendment
1.	11 (a)	<b>Airway Monitor:</b> Anaesthesia Ventilator display Monitor should be a multi-parameter module with a minimum 8 inches colour TFT display with 8 channels.	<b>Airway Monitor:</b> Anaesthesia Ventilator display Monitor should be a multi-parameter module with a minimum 8 inches colour TFT display
2	11 (c)	<b>Airway Monitor:</b> Should have 24 hours graphical and numerical trend with split screen facility of all parameters with at least 15 critical alarms summary.	This point has been deleted.
3	13	Not Mentioned	<b>Patient Monitor:</b> Should have 24 hours graphical and numerical trend with split screen facility of all parameters with at least 15 critical alarms summary.

The last date of submission of the tender has been extended till **01.06.2022** up to **02:00 PM** and the same will be opened on the same day at **3:00 PM**.

The other specifications and terms & conditions of the NIT shall remain unchanged.

*13/5/22*

Senior Administrative Officer  
Dr. B. Borooah Cancer Institute  
Guwahati – 16

Copy for information to:

1. BBCI website [www.bbcionline.org](http://www.bbcionline.org)
2. Notice Board, BBCI
3. Concerned file

*13/5/22*

Senior Administrative Officer  
Dr. B. Borooah Cancer Institute  
Guwahati – 16