

Office of the Director

**DR. BHUBANESWAR BOROAH CANCER INSTITUTE**

A grant-in-aid institute of Department of Atomic Energy, Govt. of India

And a unit of Tata Memorial Centre (Mumbai)

Gopinath Nagar, Guwahati- 781016

\*\*\*\*\*

No- BBCI-TMC/Equip- 442/ 2718 /2022

Dated: 17-06-2022

**CORRIGENDUM**

The technical specification for the supply, installation & commissioning of **Hopkin's System for Head & Neck Examination** at Dr. B. Borooah Cancer Institute against NIT No.: **BBCI-TMC/ Equip-442 / NIT/ 2072 / 2022 dated 30/05/2022** has been revised. The revised technical specification for the above mentioned equipment shall be read as follows:

Sl. No	Tender Specification as published	Amendment
B	<b>Medical Full HD Endoscope camera</b> 1. The original import CMOS chip, digital signal	<b>Medical Full HD Endoscope camera</b> 1. The original import CMOS/CCD chip, digital signal
E	<b>Rigid Tele-Laryngoscope:</b> 1. 0°C (>4 mm) 2. 70°C (>6 mm)	<b>Rigid Tele-Laryngoscope:</b> 1. 70°C (>6 mm)
		The quoted item should be USFDA & CE approved.

The last date of submission of the tender has been extended till **05.07.2022 up to 2:00 PM** and will be opened on the same day at **3:00 PM**.

The other specifications and Terms & Conditions of the NIT shall remain unchanged.

13/9 17/06/22

Sr. Administrative Officer  
Dr. B. Borooah Cancer Institute  
Guwahati - 16

Copy for information to:

1. The Assam Tribune, Guwahati  
(for publication through APS Advertising Agency)
2. BBCI website [www.bbcionline.org](http://www.bbcionline.org)
3. Notice Board, BBCI
4. Concerned file

13/9 17/06/22

Sr. Administrative Officer  
Dr. B. Borooah Cancer Institute  
Guwahati - 16