

DR. B. BOROOAH CANCER INSTITUTE

A Grant- In- Aid Institute of Department of Atomic Energy. Govt. of India

And a unit of Tata Memorial Centre, Mumbai

Gopinath Nagar, Guwahati-781016

GST No. : 18AAALB0017E1ZW

No. BBCI-TMC/Equip-441/Store/3406/2021/Corrigendum -1

Dated: 18/10/21

CORRIGENDUM

The technical specifications for item number 1. **Blood Bank Refrigerator 600 Ltr** (2-6 degree Celsius) against NIT no. BBCI-TMC/Equip-441/Store/3178/2021 dated: 28-09-2021 shall be read as follows:

Sr. No.	Sr. No. of Technical Specification	As published in Tender Document	Amendment
1	01	Storage capacity should be 360nos. of 450ml blood bags/510 nos. of 350 ml blood bags or better.	Storage capacity should have capacity to store up to 350 blood bags of 350/450ml with Internal volume approx. 500-600 liters.
2	13. d	Should have following audio-visual alarms to alert any abnormal conditions as follows: d. Heater ON – Visual	This point has been deleted
3	13. i	Should have following audio-visual alarms to alert any abnormal conditions as follows: i. Battery Low - Visual	This point has been deleted
4	21	The compressor should be placed on top of the cabinet to prevent dust accumulation and ensuring a maintenance free compressor life.	This point has been deleted

The technical specifications for item number 2. **-20 Degree Celsius Deep Freeze** against NIT no. BBCI-TMC/Equip-441/Store/3178/2021 dated: 28-09-2021 shall be read as follows:

Sr. No.	Sr. No. of Technical Specification	As published in Tender Document	Amendment
1	NIL	Not Mentioned	Laboratory deep freezer should be of more than 300 and upto 400 liters.

The other terms and condition of the NIT shall remain unchanged.


18/10/21

Dy. Director (Admin.)
Dr. B. Borooah Cancer Institute
Guwahati- 16

Copy To:

1. Tender Section of BBCI website www.bbcionline.org
2. Notice Board, BBCI
3. Concerned File